


DEXTER LifeSep™

LifeSep™ 1.5S

Biomagnetic Separator Device for 1.5mL Tubes



Dexter's series of magnetic separators for single tubes are durable and easy-to-use.

For use in cell sorting, RNA and DNA isolation and purification of biomolecules, these separation devices offer fast and easy aspiration in part because of our innovative Neodymium-Iron-Boron (NdFeB) magnet configuration.

By placing the tube in to the separator, our unique magnet configuration quickly draws superparamagnetic particles away from the solution. Once separation is complete, and with the particles held firmly to the wall and off the bottom of the tube, the supernatant can be removed.

Our LifeSep™ technology is simple and easy to use with the most established protocols.

- NdFeB magnet configuration for optimal results
- Open-faced design for clear visibility
- Rapid separation of micro- and nanometer beads
- Easy aspiration or decanting
- Superior particle retention
- Removable base for easy cleaning
- Durable and easy to use




Order online at dxtmagnetics.com or call (847) 956-1140



DEXTER MAGNETIC TECHNOLOGIES

801 Innovation Dr.
Elk Grove Village Illinois 60007 www.dxttermag.com

Siemensstraße 4
Freiburg, Germany 79108 www.dxttermag.eu



DEXTER LifeSep™

LifeSep™ 1.5S

LifeSep™ 1.5S Product Characteristics

PART NO.	VOLUME	VESSEL DIAMETER (mm)	HEIGHT (mm)	WIDTH (mm)	DEPTH (mm)	WEIGHT (kg)	NALGENE NUNC BOTTLE	DOW TUBE
2501009	1.5mL	10.5	42	49	60.5	0.1	-	3620

CARE:

The LifeSep™ 1.5S can be cleaned with mild detergent but should not be submerged in aqueous solutions.

CAUTION - STRONG MAGNETIC FIELD:

Separators should be handled with care and stored at a safe distance from magnetic media including computer hard drives, disks and credit cards.

SAFETY GUIDELINES FOR LIFESEP™ ASSEMBLIES :

The LifeSep™ 1.5S device achieves fast separation of magnetic particles through the use of powerful permanent magnets. These magnets use strong magnetic fields with very high gradients.

Although the device has been designed to focus the strongest magnetic field within the separation cavity, some magnetic field will inevitably leak outside the separation cavity. Therefore, extreme precaution should be taken when handling or otherwise being in close proximity to the LifeSep™ 1.5S device because of the following risks:

- The device poses a health and safety risk to people with pacemakers, implantable cardioverter defibrillators, and other implanted medical devices because they may interfere with the operation of those devices and/or cause them to malfunction, possibly resulting in injury or death.
- The device poses a safety risk when handling it because it can pinch fingers, hands, or other body parts that get between the magnet and another magnet or between the magnet and metal, possibly resulting in injury.
- The device poses a safety risk because it strongly attracts steel and other materials containing iron and nickel, such as tools, lab instruments, and personal items such as key chains and belt buckles. These items can be pulled forcefully to the device and can cause injury to someone holding the device or to someone who is near the device. Also, objects pulled to the device can injure the device itself.
- The device poses a risk of damage to electronic equipment, such as displays and monitors, damage to or disruption of data stored on magnetic media, such as hard drives and credit cards, and damage to magnetic recording devices. The device can also damage delicate instruments and watches.


WARNING: Because of these risks :

- No one with pacemakers or other medical implants should handle or be near the LifeSep™ 1.5S device
- No one with steel implants, such as hip or knee joint replacement implants, should permit the device to be near the implant
- All loose ferrous objects should be mechanically restrained to prevent being pulled into the device
- Remove watches and remove jewelry containing iron, steel, or nickel before handling or being near the device
- Remove all credit cards from the vicinity of the device
- Remove all electronic equipment and magnetic storage devices or media from the vicinity of the device
- Avoid having a hand or fingers between the device and any material containing iron, steel, or nickel and watches.

ABOUT LIFESEP™ :

Dexter Magnetic Technologies is the premier provider of biomagnetic separation devices with its patented LifeSep™ technology. In the most demanding process requirements, from nano-sized particle capture to continuous flow separation, we supply the technology for increased productivity.

Our technical focus enables us to understand protocols and bead separation systems. For new and existing products, we can collaborate on methods to improve rate of magnetic separation, improve wall retention, reduce cycle times and maximize yield. From single tube devices to tray units and automated solutions, we are your source for magnet expertise in separators.



DEXTER MAGNETIC TECHNOLOGIES

801 Innovation Dr.
Elk Grove Village Illinois 60007 www.dextermag.com

Siemensstraße 4
Freiburg, Germany 79108 www.dextermag.eu